



## Application

- Rapid and easy calibration of vibration transducers for acceleration, velocity and displacement
- Calibration and fault detection of vibration measuring systems
- Calibration of low frequency equipment for whole-body vibration to ISO 8041 and building vibration to DIN 4150-3

## Properties


- Load independent vibration level for test objects weighing up to 500 g
- Vibration amplitude selectable in five steps from 1 to 20 m/s<sup>2</sup>
- Quartz controlled vibration frequency selectable in seven steps from 15.92 Hz to 1280 Hz
- Display shows frequency, magnitude, error in percent and calibration date
- Rugged design
- Mains buffered battery operation for laboratory and field use
- Calibration to ISO 16062-44 with factory calibration certificate
- Upon request, we also offer DAkkS-accredited calibration with traceability



Manfred Weber

**Metra Mess- und Frequenztechnik in Radebeul e.K.**

TOTLEY bv  
Rollebeekstraat 14  
1650 BEERSEL - BELGIUM  
email : sales@totley.eu  
web : www.totley.eu  
Tel : +32 2 377 90 36



Dynamic traceable calibration of sound and vibration instrumentation and transducers, Consulting, Management, Engineering and Total Solutions provider in Electronic Instrumentation

## Technical Data

### Shaker System

Vibration frequency	15,92	40	80	159,2	320	640	1280	Hz
Vibration acceleration	1	1	1	1	1	1	1	m/s <sup>2</sup>
	2	2	2	2	2	2	2	m/s <sup>2</sup>
		5	5	5	5	5	5	m/s <sup>2</sup>
			10	10	10	10	10	m/s <sup>2</sup>
				20	20	20	20	m/s <sup>2</sup>
Vibration velocity	10			1				mm/s
	20			2				mm/s
				5				mm/s
				10				mm/s
				20				mm/s
Vibration displacement	100			1				µm
	200			2				µm
				5				µm
				10				µm
				20				µm
Weight of test object, 1 m/s <sup>2</sup>	500	500	500	500	500	500	500	g
Weight of test object, 2 m/s <sup>2</sup>	500	500	500	500	500	500	500	g
Weight of test object, 5 m/s <sup>2</sup>		500	500	500	500	500	500	g
Weight of test object, 10 m/s <sup>2</sup>			500	500	500	400	200	g
Weight of test object, 20 m/s <sup>2</sup>				250	200	100	50	g
Amplitude error, max.	±3 (0 – 40 °C)							%
	±5 (-10 - 55 °C)							%
Frequency error, max.	±0,05							%
Transverse vibration	<10	<10	<10	<10	<20	<20	<10	%
Total Harmonic Distortion (THD)	<1	<5	<1	<1	<1	<1	<1	%
Sensor mounting	M5 tapped hole (90° ± 1°; 7mm deep), magnet							
Level indication	error percent display and acoustic signal							

### Connections

Grounding connection	Banana socket 4 mm
----------------------	--------------------

### Power Supply

Battery	built-in NiMH battery pack; 7.2 V / 1.6 Ah							
Charge socket	Circular power connector to DIN 45323 (5.5 mm / 2.2 mm)							
Operating time per battery charge	5 (with 100 g weight)							h
Accumulator charging time	4							h
Charging voltage	11 – 18							V
Charging current < 1 A	<1							A
Automatic switch off	1 – 30, selectable							min

### Case Data

Dimensions without connectors	100 x 120 x 100 (W x H x D)	mm
Case material	Aluminum	
Weight	2,2	kg
Operating temperature range	-10 to 55 (95 % rel. humidity without condensation)	°C

### Scope of delivery


Plastic carrying case  
 PS1600 Mains plug adapter 100 – 240 VAC; 12 VDD; <1600 mA  
 Thread adapters M3 / M5 / M8 / 1/4"-28 / UNF 10-32

### Notice

The standard delivery includes a factory calibration sheet.  
 This is a non-accredited measurement/calibration and consequently not covered by EA MLA. On request, we offer a DIN EN ISO/IEC 17025:2018 accredited calibration of the measurand acceleration..



TOTLEY bv  
 Rollebeekstraat 14  
 1650 BEERSEL - BELGIUM  
 email : sales@totley.eu  
 web : www.totley.eu  
 Tel : +32 2 377 90 36



Dynamic traceable calibration of sound and vibration instrumentation and transducers, Consulting, Management, Engineering and Total Solutions provider in Electronic Instrumentation

TVA- BTW : BE0834.841.485 - RPR Brussel - Bank : 001-6384896-44 IBAN : BE55 0016 3848 9644 SWIFT : GEBABEBB